

CE DECLARATION OF CONFORMITY

Part numbers: RT2951 (10 cassettes), RT2952 (25 cassettes)
Product name: AMP Rapid Test SARS-CoV-2 Ag
IVD-classification acc. to directive 98/79/EC: other IVD

We, AMEDA Labordiagnostik GmbH, Krenngasse 12, 8010 Graz, Austria, declare under sole responsibility that the products described above are in compliance with Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on in-vitro diagnostic medical devices and that the following harmonized standards have been resp. are applied in development, design and manufacturing.

Instructions for Use	EN ISO 18113-2:2011	In vitro diagnostic medical devices – Information supplied by the manufacturer (labelling) – Part 2: In vitro diagnostic reagents for professional use
Performance	EN 13612:2002	Performance evaluation of IVD medical devices
Stability	EN ISO 23640:2015	In vitro diagnostic medical devices – Evaluation of stability of in vitro diagnostic reagents
Symbols	EN ISO 15223-1:2016	Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements
Risk analysis	EN ISO 14971:2012	Medical devices - Application of risk management to medical devices

The products are CE marked.

Graz - 13.08.2020


AMEDA Labordiagnostik GmbH
Krenngasse 12 - 8010 Graz - Austria
Tel. +43 - 316 - 69 80 69. Fax ext. - 12


Gerald Herfort

COO

AMEDA Labordiagnostik GmbH